



Nakawala, H., De Momi, E., Tzemanaki, A., Dogramadzi, S., Russo, A., Catellani, M., Bianchi, R., De Cobelli, O., Sideridis, A., Papacostas, E., Koupparis, A., Rowe, E., Persad, R. A., Ascione, R., & Ferrigno, G. (2019). Requirements elicitation for robotic and computer-assisted minimally invasive surgery. *International Journal of Advanced Robotic Systems*, 16(4), 1 - 32.
<https://doi.org/10.1177/1729881419865805>

Publisher's PDF, also known as Version of record

License (if available):
CC BY

Link to published version (if available):
[10.1177/1729881419865805](https://doi.org/10.1177/1729881419865805)

[Link to publication record in Explore Bristol Research](#)
PDF-document

This is the final published version of the article (version of record). It first appeared online via SAGE at <https://journals.sagepub.com/doi/full/10.1177/1729881419865805>. Please refer to any applicable terms of use of the publisher.

University of Bristol - Explore Bristol Research

General rights

This document is made available in accordance with publisher policies. Please cite only the published version using the reference above. Full terms of use are available:
<http://www.bristol.ac.uk/red/research-policy/pure/user-guides/ebr-terms/>

Requirements elicitation for robotic and computer-assisted minimally invasive surgery

Hirenkumar Nakawala¹ , Elena De Momi¹, Antonia Tzemanaki², Sanja Dogramadzi², Andrea Russo³, Michele Catellani³, Roberto Bianchi³, Ottavio De Cobelli^{3,4}, Aristotelis Sideridis⁵, Emmanuel Papacostas⁵, Anthony Koupparis⁶, Edward Rowe⁶, Raj Persad⁶, Raimondo Ascione⁷ and Giancarlo Ferrigno¹

Abstract

The robotic surgical systems and computer-assisted technologies market has seen impressive growth over the last decades, but uptake by end-users is still scarce. The purpose of this article is to provide a comprehensive and informed list of the end-user requirements for the development of new generation robot- and computer-assisted surgical systems and the methodology for eliciting them. The requirements were elicited, in the frame of the EU project SMARTsurg, by conducting interviews on use cases of chosen urology, cardiovascular and orthopaedics procedures, tailored to provide clinical foundations for scientific and technical developments. The structured interviews resulted in detailed requirement specifications which are ranked according to their priorities. Paradigmatic surgical scenarios support the use cases.

Keywords

Robot-assisted surgery, computer-assisted interventions, user requirement specifications

Date received: 28 January 2018; accepted: 4 June 2019

Topic: Medical Robotics

Topic Editor: Arianna Menciassi

Associate Editor: Giuseppe Carbone

Introduction

Computer-assisted surgery (CAS) or robotically assisted surgery (RAS) relies on specially devised machines and equipment to enhance a part of or an entire surgical procedure. Surgical robotic systems are a combination of robotic and imaging equipment, surgical accessories, software and services that assist surgeons in performing various types of minimally invasive surgery (MIS) (e.g. gynaecological, cardiovascular, neurological, urological and orthopaedic). Robotic systems facilitate dexterous teleoperation of surgical tools, with a view to reduce invasiveness and to improve efficacy and precision by miniaturization, improved stability, natural coordination, enhanced surgeon ergonomics and increased dexterity. Additional benefits may include

¹ Department of Computer Science, University of Verona, Verona, Italy

² Bristol Robotics Laboratory, University of the West of England, Bristol, UK

³ Division of Urology, European Institute of Oncology, Milano, Italy

⁴ Department of Oncology and Hematology-Oncology University of Milan, Milan, Italy

⁵ TheMIS Orthopaedics Centre, Thessaloniki, Greece

⁶ Bristol Urological Institute, North Bristol NHS Trust, Bristol, UK

⁷ Translational Biomedical Research Centre and Bristol Heart Institute, University of Bristol, Bristol, UK

Corresponding author:

Hirenkumar Nakawala, Department of Computer Science, University of Verona, Via Strada le Grazie 15, 37134 Verona, Italy.

Email: hirenkumarchandrakant.nakawala@univr.it



Creative Commons CC BY: This article is distributed under the terms of the Creative Commons Attribution 4.0 License

(<http://www.creativecommons.org/licenses/by/4.0/>) which permits any use, reproduction and distribution of the work without further permission provided the original work is attributed as specified on the SAGE and Open Access pages (<https://us.sagepub.com/en-us/nam/open-access-at-sage>).

reducing post-surgical complications, blood loss, post-operative pain, better tissue healing as well as faster recovery and reduced hospital costs compared to conventional surgery.^{1,2} In addition, robotic manipulators help surgeons to work on less accessible parts of the body via small ports. For example, robotic surgery has been extensively used to enhance the excision of prostate, affected by cancer, given the difficulty in accessing the prostate in conventional laparoscopic surgery, which requires extensive experience.² Collectively, the characteristics of the robotic systems aim to reduce surgical invasiveness and enhance patient outcomes.

The global medical robots market is expected to reach US\$12.80 billion by 2021 from US\$4.90 billion in 2016, growing at a Compound Annual Growth Rate of 21.1% (<http://www.marketsandmarkets.com/PressReleases/medical-robotic-systems.asp> (accessed 20 November 2017)). Key players in the global medical robots market include Intuitive Surgical Inc. (Sunnyvale, CA, USA), Stryker Corporation (Kalamazoo, MI, USA), Mazor Robotics Ltd (Israel), Hocoma AG (Switzerland), Hansen Medical Inc. (Mountain View, CA, USA), Accuray Incorporated (Sunnyvale, CA, USA), Omnicell, Inc. (Mountain View, CA, USA), Ekso Bionics Holdings, Inc. (Richmond, CA, USA), ARxIUM (Buffalo Grove, IL, USA) and Kirby Lester LLC (Lake Forest, IL, USA) (<http://www.marketsandmarkets.com/PressReleases/medical-robotic-systems.asp> (accessed 20 November 2017)). To date, the race to develop, acquire and incorporate this emerging biomedical technology has been primarily driven by the market and the industry. While several robot-assisted MIS (RAMIS) systems have gone through feasibility trials for clinical validation, regulatory approvals and commercialization, their actual use is still scarce compared to conventional surgical procedures. Global adoption of these systems in clinical practice is still sporadic³ except for the da Vinci surgical system (Intuitive Surgical Inc., USA) which has played a major role in RAMIS in the last two decades. Although the use of robotic systems has been tested in several surgical subspecialties, that is, cardiovascular, thoracic, urological, gynaecological, paediatric, and general surgery,³ surgeons in these specialties are still not inclined to use these technologies. This lack of market penetration may have been triggered by poor training programmes, excessive cost, need for reconfiguration of conventional surgical theatres as well as reported suboptimal results in some surgical areas.⁴ The U.S. Food and Drug Administration (FDA, Maryland, USA) has reported a marked increase in adverse events associated with robotic surgery during 2006–2013, and this may have triggered more caution in the surgical field while prompting refocus for the development of more advanced and safer devices.⁴ Surveys with surgeons suggest that inclination to the technologies is often associated with the gold standard for a surgical procedure at a given time, which is connected to the statistical success of a specific technique and indirectly to its cost.⁵ Furthermore,

it has been reported that only 3% of the RAMIS cystectomies in the United States were completed without converting to the open technique.⁶ The slow adoption is perceived to be increased difficulty due to the handling of sensitive structures as well as working in a confined space.

The Society of American Gastrointestinal and Endoscopic Surgeons has initiated work on finding a consensus on robotic surgery including guidelines for training and credentials, indicating goals on instrumentation, visualization, integration and simulation.⁷ Efforts have been carried out to assure that patient safety is the top priority when envisaging a new robotic solution to be used to enhance surgical performance.⁸ In a previous study on decision-making and assistive robotic technologies in surgery,⁹ a qualitative analysis was performed to evaluate four critical characteristics of surgical assistance systems, that is, situational awareness, lack of tactile feedback, immersion and impact of ergonomics. This study also provides surgeons' requirements that are essential in designing more advanced robotic systems. However, the study was only focused on one surgical specialty, that is, colorectal surgery and was constrained to the context of decision-making. Bonfè et al.¹⁰ used the requirements engineering methodology to collect surgeons' requirements for a software-intensive and intelligent surgical robot on three chosen actions, that is, needle insertion for ablating procedure, laparotomy and suturing a wound. A group of experts was interviewed to obtain the goal model, which was used to express structural constraints and behaviours in a software system. In the study by Stollnberger et al.,¹¹ a group of stakeholders, that is, doctors, patients and assistants, was interviewed, and the feedback was collected for the development of a robotic medical system which is able to conduct ultrasonography and physical examination remotely. The latter study was focused on simple use cases, such as suturing a wound. Facilitating factors and barriers for adopting robotic systems among health-care professionals have also been investigated.³ After semi-structured interviews, one of the main facilitator factors for adopting a robotic system was 'Perceived Usefulness'. 'Perceived Usefulness' regards functions with the robot – better visualization, increased precision, better dexterity, elimination of hand tremor, better suturing, better instrumentation, better angle of placement, easier access and better ergonomics. A new RAS system is expected to have the perceived usefulness. Relevant shortcomings of the current system functionality could be elicited by analysing end-user requirements. Titan Medical Group (Canada) has included a process focused on a clear and limited set of customer-centric requirements in the development of a single-port robotic surgical system.¹² As presented,¹³ the development of first-generation robotic systems was generally focused on early stage developments. With no following through, the systems lack capturing in the final product of critical later stage developmental aspects. Also, earlier studies were limited in duration and




Figure 1. Definitions are shown in italics, for example, ‘Surgical phase’. The orange arrows show the hierarchical relation between these definitions, for example, ‘Surgical step’ is a part of ‘Surgical phase’. The figure shows an example of ‘Tumour Excision’ phase of RAPN. RAPN: robot-assisted partial nephrectomy.

showed little involvement from the end-users. It is consequently debatable whether they cover their actual needs. A qualitative study by Aaltonen and Wahlström¹⁴ was focused on three aspects: enhancing surgical operation outcome, user experience and learning. The study selected technological solution concepts based on a technology review and an ethnographic study. It is acknowledged by the authors that the study does not provide a complete list of user requirements nor it facilitates meaningful discussions. Also, the ethnography study is only useful to address the contextual factors such as usability and investigating collaborative work settings,¹⁵ for example, in the case of software requirements elicitation too.

The purpose of this qualitative study was to gather the end-user requirements to design an improved surgical system, that is, within the SMARTsurg project (<http://smart-surg-project.eu/> (accessed 20 November 2017); <https://www.youtube.com/watch?v=AgPkMSqXRfs&t=15s> (accessed 20 October 2018)), by determining the barriers of the methods and systems currently used and what advancements are needed specifically for the user interface (master), surgical instruments, vision and features such as haptic feedback and so on. The SMARTsurg project aims at developing an advanced system for RAMIS, focusing on reducing the surgeon’s cognitive load related to the system’s operation to enhance shorter training time, while delivering more accuracy, safety, reduced procedure time and expanded surgical applicability. To this end, the project aims to design and develop a wearable interface for a surgical system using (a) highly dexterous surgical instruments, (b) wearable hand exoskeletons with haptic feedback and (c) wearable smart glasses for augmented reality guidance of the surgeon based on the real-time three-dimensional (3D) reconstruction of the surgical field. Previous work included investigations on a wearable system of three-finger hand exoskeletons and anthropomorphic

three-finger gripper.¹⁶ High dependability will be achieved by utilizing real-time dynamic, active constraints (ACs) to the instruments’ motion in order to restrict it to the safe regions. SMARTsurg developments will employ a user-centred approach for efficient technology adoption and commercialization. This will be achieved using short prototyping and testing cycles supported by focused end-user and commercial requirements.

We followed a qualitative data analysis approach¹⁷ for gathering the multi-user specific requirements, where urologists, cardiac and orthopaedic surgeons were involved considering the design of SMARTsurg system and targeted procedures. This article aims to illustrate the results of the requirements elicitation derived for the urology, cardiovascular and orthopaedic use cases, which would be accounted in the system design and implementation process. Once the current standardized workflow was defined using the controlled vocabulary in the graphical format, elicited requirements have been prioritized according to surgeons’ needs. Also, the conceived surgical scenarios by mapping requirements to surgical workflow are herewith presented.

Methods

Definitions and controlled vocabularies

In order to provide a standard methodology for the surgical intervention descriptions, for example, as shown in Figure 1, the consensus was reached among the SMARTsurg consortium on the following terminology:

- (i) *Surgical phases* indicate a collection of surgical activities, which need to be performed in a particular sequence to accomplish the surgical workflow. The aim of each phase is to reach/target the principal surgical site. For example, in the

- 'Tumour excision' phase of robot-assisted partial nephrectomy (RAPN), which involves removal of kidney tumour, the surgeon first identifies the site for tumour by cutting Gerota's fascia, then he/she does the markings on the kidney capsule to expose the tumour area for resection.
- (ii) *Surgical steps* are the tasks required to accomplish the phases of the procedure. Each step consists of a specific action, anatomical locations and instruments. For example, during 'Tumour excision' (phase), the surgeon does the 'clamping' (step) of the 'renal artery' (anatomical location) by 'clamp' (action) through the 'Laparoscopic Bulldog clamp' (instrument). Sometimes, the steps correspond to the same linguistic meaning, where the phases consist of only one step. For example, 'Bowel mobilization' phase has only one step, that is, 'mobilization'.
 - (iii) *Surgical actions* are carried out by surgical instruments in surgical steps, for example, 'cortical suturing' (step) by the 'large Needle Driver' (instrument) to 'suture' (action) the 'kidney' (anatomy) during the repair of the kidney, 'renorrhaphy' (phase), at the end of the procedure.
 - (iv) *Surgical instruments* are used during a single step of the surgery. Surgical instruments indicate robotic instruments, for example, 'fenestrated Bipolar', in left and right robotic arm. Instruments, for example, 'laparoscopic Bulldog', used by assistant surgeons are also considered.
 - (v) *Anatomical locations* are the anatomical landmarks, for example, 'gerotas Fascia', on which an instrument performs a surgical action.

Use cases

The use cases were selected to contextualize the focused requirement elicitation process and to elicit application scenarios that specify targeted phases and steps, which collectively form workflows, during implementation and evaluation of system's components. The identified use cases are as follows:

- (1) Orthopaedic surgery
 - a. Robot-assisted partial lateral meniscectomy (RaPLM): A meniscus tear is a common knee joint injury. RaPLM is the surgical removal of all or part of a torn meniscus.
 - b. Robot-assisted repair of lateral meniscus tear (RaLMR): A meniscus tear, if not severe, RaLMR is performed to repair the part of a torn meniscus.
- (2) Urological surgery
 - a. Robot-assisted radical cystectomy (RARC) and intracorporeal reconstruction with ileal conduit or orthotopic neobladder: RARC is performed to remove the cancerous bladder. Here, intracorporeal reconstruction with ileal conduit or orthotopic neobladder indicates the reconstruction of urinary diversion to urethra after removal of the bladder.
 - b. Robot-assisted radical prostatectomy (RARP): RARP regards prostate cancer removal, where the entire prostate gland along with some of the surrounding tissue is removed.
 - c. RAPN: RAPN regards kidney tumour removal, where a tumorous portion of the kidney is removed.
- (3) Cardiovascular surgery
 - a. Mitral valve (MV) leaflet repair (MV mending surgery): MV mending surgery is done when the MV is too loose or severely leaking. MV replacement with an artificial valve is done when the MV is too calcified and not opening well.
 - b. Coronary artery bypass grafting (CABG): CABG is advised for a selected group of patients with significant narrowing or blockage of the coronary artery.

The detailed procedural workflows (consisting of phases, steps and instruments) of the use cases are outlined in Appendix 1. Procedural definitions and use cases workflows were confirmed after a consensus within the clinical partners of the consortium. Each surgical workflow includes workflow entities, for example, 'Phase', 'Steps' and 'Instruments'.

User requirement collection and data analysis

We interviewed non-expert and expert surgeons in the surgical specialties. They expressed their views on potential barriers, limitations and improvements of current surgical systems for CAS and RAMIS. We conducted a total of 29 interviews. As per the breakdown of specialties, we interviewed 6 orthopaedic surgeons (1 senior, 3 mid-careers and 2 juniors), 17 urologists (7 seniors, 3 mid-careers and 7 juniors) and 6 cardiac surgeons (4 seniors, 1 mid-career and 1 junior). The mean age of orthopaedic surgeons, urologists and cardiac surgeons was 41, 43 and 39.2 years, respectively. Urologists were from Italy and the United Kingdom. Orthopaedic surgeons were from Greece, while the cardiac surgeons were from the United Kingdom and Greece. The interviewees are all male surgeons from Europe. However, in a span of 1 year, this was the maximum achievable number of surgeons within the specialties with a higher gender ratio, for example, urology, cardiovascular surgery and orthopaedic surgery. For example, the female surgeons count for only 8% in cardiac surgery, 10% in urology and 6% in orthopaedics in the United Kingdom.¹⁸ In Italy, less than 10% of urologists are female. Worldwide statistics of urologists' gender is difficult to find, but in most countries,

Table 1. Interviewed surgeons' experience with open surgery, MIS and RAMIS.

| Specialty | Open surgery (experience in years) | MIS (experience in years) | RAMIS (experience in years) |
|-----------------|--|---|---|
| Orthopaedics | More than 7 – 2 surgeons; 5 to 6 – 2 surgeons; 3 to 4 – 1 surgeon; 1 to 2 – 0; less than 1 – 0; NULL – 1 surgeon | More than 7 – 2 surgeons; 5 to 6 – 1 surgeons; 3 to 4 – 1 surgeon; 1 to 2 – 2 surgeons; less than 1 – 0 | Less than 1 – 3 surgeons; NULL – 3 surgeons |
| Urology | More than 7 – 7 surgeons; 5 to 6 – 3 surgeons; 3 to 4 – 5 surgeons; 1 to 2 – 0; less than 1 – 2 surgeons | More than 7 – 4 surgeons; 5 to 6 – 4 surgeons; 3 to 4 – 2 surgeons; 1 to 2 – 3 surgeons; less than 1 – 3 surgeons; NULL – 1 surgeon | More than 7 – 4 surgeons; 5 to 6 – 3 surgeons; 3 to 4 – 5 surgeons; 1 to 2 – 3 surgeons; less than 1 – 2 surgeons |
| Cardiac surgery | More than 7 – 6 surgeons | More than 7 – 1 surgeon; 5 to 6 – 1 surgeon; less than 1 – 3 surgeons; NULL – 1 surgeon | Less than 1 – 3 surgeons; NULL – 3 surgeons |

MIS: minimally invasive surgery; RAMIS: robot-assisted MIS.

majority of urologists are male, for example, in Japan 95% and in the United States about 93%.¹⁹ As for surgeons' skill levels, we interviewed 12 seniors, 7 mid-careers and 10 junior surgeons. A detailed information on the surgeon's experience with open, MIS and RAMIS is shown in Table 1.

Expertise level was determined by surgeons themselves. We interviewed surgeons with different levels of experience, where the questions were particularly designed consisting of understanding the limitations of open and laparoscopic surgeries as well, considering the lack of RAMIS experience of orthopaedics and cardiovascular surgeons as compared to urologists. Moreover, during the interviews, the surgeons were provided with the printed slides of surgical robotics technologies (see the Online Supplementary Material), which were explained to them to understand how certain technologies would be helpful to improve proposed use cases. The slides consist of reported examples of physical hardware blocks, for example, master-slave design, vision components and so on which explained to be not taken as the foreseen solutions, but only as suggestions that will be adaptable to the actual surgical requirements. The surgeons allowed to freely think of their own application and to imagine the complexity they need without sticking to reported images. Finally, after the interview, a slide showing the prospective SMARTsurg system was explained and specific questions related to the system were asked, for example, 'How do you expect a system like SMARTsurg will improve in new surgeons training?'

Structured interviews were conducted either face-to-face, telephone or via computer call. In all cases, interviews were recorded in audio format as raw data. Participants gave written informed consent, and the data collection procedure was approved by Politecnico di Milano Ethical committee (opinion n. 512017). Similarly, the University of the West of England interviews and data collection were done in accordance with the recommendations of the University's policy on research ethics, approved by the Faculty of Environment and Technology Research Ethics Committee (UWE REC REF No: FET.17.04.038).

Two types of questions were used during the interviews

- (1) 'Open-ended' questions, where surgeons expressed their opinions in the descriptive form; and
- (2) 'Close-ended' questions, where surgeons gave the answers in the form of Yes/No or surgeons expressed the answers by selecting one or more options (categories/concepts).

After the interviews, recordings were transcribed, and the data were subsequently organized. Answers were grouped for each question in the questionnaire. We assigned each surgeon an ID, that is, the first letter of each specialty followed by the user number, for example, O1, O2 and so on for orthopaedic surgeons; U1, U2 and so on for urologists; and C1, C2 and so on for cardiac surgeons.

The first analysis was conducted employing 'within-case analysis'¹⁷ method, where surgeons' responses for individual surgical case study were explored in detail, as a standalone entity, to discern the patterns revealed in the individual interviews (e.g. 'within-case analysis' of collected interview data of orthopaedic surgeons, urologists and cardiac surgeons separately). The 'within-case analysis' was used to identify common categories/concepts from each surgical use case. To construct the categories, we did manual open coding.²⁰ A code is a word, phrase or sentence that represents aspect(s) of the data or captures essence or features of the data. The purpose of coding is to reduce the data into meaningful segments and assign names (codes) to those segments. The names of categories were defined by the domain expert in surgical robotics, or by participant's exact words or the literature sources relevant to the study. Categories are related to (1) the phenomenon under study; (2) the contextual, intervening-structural and causal conditions; (3) the actions to handle the phenomenon; and (4) consequences of actions and interactions related to phenomenon.²⁰ For example, in the sentence,

"Surgeons are familiar with the use of instruments. Generally, there are problems with the tissues e.g. thin meniscus. We may

Table 2. Priority level and associated scores for the elicited user requirements.

| No. | Priority | Score |
|-----|-------------|-------|
| 1 | High | 5 |
| 2 | Medium-high | 4 |
| 3 | Medium | 3 |
| 4 | Medium-low | 2 |
| 5 | Low | 1 |

need smaller instruments. Current instrument diameter is approximately 2 cm.”

The code for this sentence is ‘small instruments to manage tissue consistency’. However, the categories are ‘anatomical problems’ and ‘small instruments’. Here ‘Anatomical problems’ is a causal factor for the requirement of small instruments. The categories define themes which are used to identify a major element of the content analysis of the text.

Further on, a disaggregation of core themes/categories, that is, ‘axial coding’ was applied to the collected information.²¹ Axial coding is the process of relating codes (categories and concepts) via a combination of inductive and deductive thinking. The grouped categories are also mapped to prospective system hardware components. There are also decision blocks, for example, ‘Are all the user requirements examined?’ regards the elicited requirements that would help eliciting essential requirements of a system, where ‘NO’ decision states reanalysing the transcripts and the defined codes, and ‘YES’ decision states the essential user requirements are met and no further analysis is required. Closed questions, which inform explicit requirements to test surgeon’s opinion on them, were analysed using the analytical approach. In such cases, we found the requirements by analysing the categorical data.

‘Across-case’ analysis and elicitation of application scenarios

To do the ‘across-case analysis’, the elicited requirements obtained using the ‘within-case analysis’ were first prioritized and scored, as shown in Table 2.

The priority levels and scores for user requirements were obtained from a consensus among the clinical partners during the SMARTsurg 1st PC Meeting (Milan, Italy, 10–11 of July 2017). After deciding the priorities and scores, each of the elicited requirements with the same categories from different specialties was grouped together.

User requirements were considered mandatory requirements, that is, for the SMARTsurg system, if total priority scores of requirements from three specialties were summed up to ≥ 14 . The non-mandatory requirements (total score ≤ 13), which include high and medium-high requirements (total score ≥ 10 but ≤ 13), were also analysed with the

same method, but these priorities may be further extracted from the application scenarios. The threshold scores on the elicited requirements were decided as a trade-off between their complexity and the project’s resources. To extract the application scenarios, elicited requirements for each specialty were mapped to the individual phases and steps of use cases considering ‘within-case’ and ‘across-case’ analysis. Information on the use case phases and steps were obtained from the use cases’ workflows. The full user requirements analysis methodology is shown in Figure 2.

Results

‘Within-case’ analysis

Appendix 2 (Tables 2A to 2C) represents, ‘within-case’ analysis of surgeons’, feedback obtained through the interviews. Multiple utterances or discussions by the same surgeon were also considered for eliciting the requirement if the meaning of the utterances was different concerning the requirements. In each table, surgeons’ feedback is reported for each feedback/requirement in the sequence of questions in user requirement questionnaire. The ‘within-case’ has identified 13, 18 and 14 different categories of elicited requirements of orthopaedics, urology and cardiac surgery use cases, respectively. For orthopaedic surgery, the category ‘anatomical problem’ was discussed seven times, more than any other category. The haptic feeling (17 times) and image quality (9 times) were predominantly discussed for urology and cardiovascular surgery use cases. In all the specialties, vision is stated as a barrier due to the small and difficult to access anatomical structures by camera. Contrary to this, the urologists, who have greater experience with the RAMIS than the other two specialist groups, discussed the need to improve the current camera systems, highlighting the limitation of camera resolutions, larger length of the camera shaft and so forth. Similarly, all surgeons also are in favour of small articulated and flexible instruments which could overcome the problem of efficiently accessing the anatomical structures, for example, MV or meniscus structure. The need for haptic feeling when suturing, dissecting and identifying the normal tissue, for example, kidney as well as the abnormal tissues, for example, tumour was also prioritized. All surgeons agreed that their postures are non-ergonomic during the surgery. Having a third tool digit is considered for tissue manipulation, for example, stabilization, repairing and suturing, as well as grasping and replicating instrument’s movements as in, for example, Castroviejo-type instruments for heart surgery. However, surgeons need graspers which could provide more force to grasp the tissue. Instrument tip swapping could be helpful in reducing infections, saving the operative time as well as replacing the task performed by assistants. For more interactive experience, urologists and cardiovascular surgeons prefer immersive stereo viewer, while orthopaedic surgeons prefer the smart




Figure 2. A flow chart of the user requirements analysis methodology.

glasses. Superimposing preoperative images to identify anatomical structures intraoperatively was favoured by all surgeons. Only the cardiovascular surgeons would like to see physiological data intraoperatively, for example, blood pressure. They also need communication with their team, where immersive stereo viewer could be helpful in order to provide a holistic view of the procedure. As an alternative sensory information, visual cues are most favourable except for cardiovascular surgeons. While orthopaedic surgeons do not need to have any ACs, urologists and cardiovascular surgeons think that this could be a helpful feature for preventing injuries of the critical structures, for example, blood vessels

but also for the surgical training. On the topic of surgical training, some surgeons prefer the existing training methodology, while some, especially urologists, would like to use these technologies, for example, haptics to improve understanding of the procedure and dexterity. The major concerns about these technologies are their costs, teleoperation for manipulating tissues (especially to orthopaedic surgeons) and patient safety, usability and improvement with respect to the currently available robotic systems, for example, synchronized simultaneous movements of slave robot arms with the patient table. A detailed information on results of 'within-case' analysis is shown in Appendix 2. Functional

Table 3. Functional requirements.

| No. | Requirements | Orthopaedics | Urology | Cardiovascular surgery |
|-----|---|--|---|---|
| 1 | Size of instruments | The diameter of new instruments should be less than 4 mm or similar to the diameter of current instruments. In case of flexible instruments, orthopaedic surgeons require three degrees of freedom | Flexible instruments should provide 360° rotational movement with a length from 0.5 cm to 1.0 cm. Pro-Grasp forcep should be less than 1.5 cm | Flexible instrument should provide 360° rotational movement. The needles, that is, with 7–0 or 9–0 Prolene sutures, should be thin |
| 2 | Field of view | It should be less than 1 cm ² to 6 cm ² | It should be less than 5 cm ² to 25 cm ² | It should be from 1.5 mm ² to 6 cm ² |
| 3 | Tolerable registration error while superimposing preoperative images | 2–3 mm during meniscus repair | 2 mm for RAPN | 0.5–1 mm for CABG; 1–2 mm for MV surgery |
| 4 | Weight of the wearable hand exoskeleton | Up to 500 gm | Up to 500 gm | 25–100 gm |
| 5 | Workspace requirements (how much workspace is needed by the surgeon to perform the procedure comfortably?) | 1.5 m ² | Surgeons perform the procedure in sitting position. Surgeons need the space for movement and safety | Surgeons generally stand in the area of 40 cm ² for 2–5 h. If operating at a robotic station, in a sitting position, the space will be that of a smart and ergonomic chair with no sitting down/back approach. |
| 6 | Resolution of images | 4K resolution (4096 × 2160) | HD resolution (1280 × 720) or UHD resolution (3840 × 2160) | 3.5–4.5× or more 3D magnification over current image resolution |
| 7 | Haptic feedback in terms of scale from 1 to 10, for example, 1 = <i>Very low haptic feedback</i> to 10 = <i>Very high haptic feedback</i> | 5–8 | 7 | For CABG, it must be 10. For MV surgery, the scale of haptic feedback should be within the range of 7–10. |

MV: mitral valve; CABG: coronary artery bypass grafting; RAPN: robot-assisted partial nephrectomy; 3D: three-dimensional; HD: high-definition; UHD: ultra high-definition.

requirements, which are elicited from the interviews, are reported in Table 3. Functional requirements could be helpful in the development of technical specifications for system components.

Across-case analysis

In Table 4, each cell is represented with the elicited mandatory requirement with its necessity in three specialties, that is, urology (U), orthopaedics (O) and cardiac surgery (C) (priority levels: 5 = high; 4 = medium-high; 3 = medium; 2 = medium-low; and 1 = low). After carrying out ‘within-case’ analysis, we elicited 13, 18 and 14 distinct categories of elicited requirements for orthopaedics, urology and cardiovascular surgery, respectively. A total of 33 user requirements have been elicited, on which the across-case analysis is done, out of which 4 requirements (e.g. superimposed preoperative images, ACs, articulated instruments and hand exoskeleton as a master system) are the mandatory requirements, that is, priority score ≥ 14 . After the ‘across-case’ analysis, application scenarios were chosen. The main reason to conduct across-case analysis was to allow for a

versatility of the systems matching the requirements in order to be used for different purposes in the hospital. Examples of application scenarios on all the use cases are reported in Table 5. The non-mandatory requirements (from no. 5 to 33) are presented in Appendix 3.

Discussion

We discuss four mandatory requirements, that is, total score ≥ 14 , namely ‘superimposed pre-operative images’, ‘Active constraints’, ‘Articulated instruments’ and ‘Master system – hand exoskeleton’. The discussion is split specialty-wise, following orthopaedics, urology and cardiovascular surgery, except for the ‘Master system – hand exoskeleton’, where the requirement was discussed considering all specialties.

Superimposed preoperative images

Orthopaedic surgeons use X-ray and magnetic resonance imaging (MRI) as preoperative images. They were not sure if it is possible to superimpose preoperative images

Table 4. ‘Across-case’ analysis – mandatory requirements.

| No. | Elicited requirement | Orthopaedics | Urology | Cardiovascular surgery | Total score |
|-----|---|---|--|---|-------------|
| 1 | Preoperative images superimposed on the intraoperative scene Details | Information on preoperative images (CT, USG, MRI) needed, for example, to minimally cut meniscus | Information on preoperative images (MRI) are needed. | Information on physiological data and preoperative images (combination of echocardiography, coronary angiography, CT scans, MRI) are needed | |
| | Scores | 5 | 5 | 5 | 15 |
| 2 | ACs Details | AC is helpful to prevent injury to rim of the meniscus, to remove only the damaged meniscus or the meniscus flaps | AC is helpful to prevent injury to damaged nerves, small or big vessels, for example, aorta, vena cava and supplementary vascularisation, for example, extra-renal artery, and during lymphadenectomy step of prostatectomy. AC may also useful for the training | AC is needed to prevent injury to vessels and nerves. For example, AC could prevent burning of the LIMA while using the cautery in CABG | |
| | Scores | 5 | 5 | 5 | 15 |
| 3 | Articulated instruments Details | Small articulated instruments are needed to work in narrow space inside the knee | Articulated instruments, at least with two articulations, are required because of small and close structures in the pelvis, for example, ridges of pubic bone or peculiar shape of a pubic bone. It is needed in complex cases such as previous multiple pelvic or abdominal procedures or pelvic adhesion | Articulated instruments are required to reach or manipulate some anatomical structures, for example, MV, ventricles behind the MV, and during the cross-clamping of aorta | |
| | Scores | 5 | 5 | 4 | 14 |
| 4 | Master interface Details | Hand exoskeleton would be a feasible option as a master interface | Hand exoskeleton would be a feasible option as a master interface | Hand exoskeleton would be a feasible option as a master interface since it also provides efficient wrist movements | |
| | Scores | 5 | 4 | 5 | 14 |

USG: ultrasonography; AC: active constraint; CABG: coronary artery bypass grafting; LIMA: left internal mammary artery; MV: mitral valve; CT: computed tomography.

because preoperative and intraoperative images often have different orientations. The preoperative images are being taken in the supine position when the knee joint is flexed during the surgery. However, there are a few landmarks that could be useful for image registration, for example, medial and lateral femur condyle, anterior cruciate ligament, trochlea and medial compartment of the tibia. Urologists use computed tomography (CT), ultrasonography and MRI as preoperative images. The preoperative and intraoperative images are always a little different for urology use cases. There is a little difference in the

parenchymal organs, for example, kidneys, but images could change for other organs, for example, the peritoneum so that the image registration could be difficult. However, superimposed images could be helpful to understand the relative positions of the organs, for example, where a tumour or ureter is. Urologists suggested landmarks that could be useful for registration, for example, vessels like the aorta, organs like spleen, lower and upper poles of the kidney during RAPN, nerves, seminal vesicles, pubic bone and the apex of the prostate during RARP, and the middle lobe of the prostate and pubic symphysis during RARC.

Table 5. Application scenarios for RAPN, RARC, RARP, RaPLM, RaLMR, MV surgery and CABG.^a

| No. | Elicited requirement | RAPN | RARC | RARP | RaPLM | RaLMR | MV surgery | CABG |
|-----|--------------------------------------|--|--|--|--|---|---|---|
| 1 | Hand exoskeleton as master interface | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps |
| 2 | Articulated instruments | – | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps | Phase 1 and steps 1.1, 1.2, 1.3, 1.4, 1.5, 1.6 and 1.7 |
| 3 | Three-fingered instrument | – | All phases and steps | All phases and steps | Phase 2 and steps 2.1, 2.2, 2.3 and 2.4; phase 4 and steps 4.1a, 4.1 and 4.2 | Phase 3 and steps 3.1, 3.2, 3.3, 3.4 and 3.5; phase 4 and steps 4.1a, 4.1b and 4.1c | All phases and steps | Phase 2 (LIMA-LAD anastomosis via left minithoracotomy) and step 2.4; phase 2 (LIMA-LAD anastomosis via closed chest robotics approach) and steps 2.2, 2.3 and 2.4 |
| 4 | High-resolution images | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps |
| 5 | 3D Images | All phases and steps | All phases and steps | All phases and steps | – | All phases and steps | All phases and steps | All phases and steps |
| 6 | Superimposing preoperative images | Phase 5 and steps 5.1 and 5.2 | – | Phase 7 and its all steps; phase 8 and its all steps | Phase 4 and steps 4.1 and 4.2 | Phase 4 and steps 4.1a, 4.1b and 4.1c | Phase 2 and steps 2.1, 2.2, 2.3 and 2.4 | Phase 1 and steps 1.1, 1.2, 1.3, 1.4, 1.5 and 1.6 |
| 7 | Smart glasses (for assistants) | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps | All phases and steps |
| 8 | ACs | Phase 1 and steps 1.2, 1.3 and 1.6; phase 5 and step 5.2 | Phase 1 and its all steps; phase 2 and its all steps | Phase 1.4 and its all steps | Phase 4 and steps 4.1 and 4.2 | Phase 4 and steps 4.1a, 4.1b and 4.1c | All phases and steps | Phase 1 and steps 1.3, 1.4 and 1.5 |
| 9 | Haptics | Phase 6 and steps 6.1, 6.3, 6.4 and 6.7 | All phases and steps | All phases and steps | Phase 3 and steps 3.3 and 3.4 | Phase 3 and steps 3.1, 3.2, 3.3, 3.4 and 3.5 | All phases and steps | Phase 1 and step 1.7; phase 2 and step 2.4 (LIMA-LAD anastomosis via left minithoracotomy); phase 2 and steps 2.2 and 2.3 (LIMA-LAD anastomosis via closed chest robotics approach) |

AC: active constraint; RAPN: robot-assisted partial nephrectomy; RARC: robot-assisted radical cystectomy; RARP: robot-assisted radical prostatectomy; RaPLM: robot-assisted partial lateral meniscectomy; RaLMR: robot-assisted repair of lateral meniscus tear; MV: mitral valve; CABG: coronary artery bypass grafting; LIMA: left internal mammary artery; LAD: left anterior descending artery; 3D: three-dimensional.

^aThe table demonstrates the use of technology (elicited requirement) with respect to the use cases of the study. Phases and steps correspond to the use cases workflow, as shown in Appendix 1. As an example, application scenarios on RAPN such as hand exoskeleton, better image quality, smart glasses (for assistants) and 3D images for visualization are needed in all the phases of RAPN. During preparation of the kidney, ACs could be used to prevent the injuries to vasculature such as the aorta or vena cava as well as organs such as the liver and the spleen. After that, during the excision of the tumour, preoperative images are superimposed to see the renal artery, while incising the renal capsule before clamping the artery. After the preoperative images are superimposed, ACs could be used to prevent the injury to renal arteries. Afterwards, haptics could be used for the closure of renal breach during suturing and pulling of the thread while doing the suturing of the kidney.

Urologists suggested that superimposed images are useful in specific surgical steps of these use cases as, for example, during the nerve-sparing in RARP or to identify a tumour during RAPN because these anatomical regions are visible on MRI. They suggested that superimposing preoperative images could also be useful to identify the enlarged lymph nodes in unusual locations. However, surgeons need ‘on and off’ functionality for this feature. In RARP, base of the prostate is clearly visible which provides the precise coordinates as well as the apex. These coordinates can be used to determine the site of lesions for the image fusion. Cardiac surgeons use a combination of echocardiography, coronary angiography, CT or MRI as preoperative images. They suggested that it is possible to superimpose preoperative images because there is not much difference between preoperative and intraoperative images for these two use cases. However, it is hard to define the landmarks in beating heart surgery. Otherwise, there are enough landmarks available, for example, appendages or great vessels such as the aorta and the apex of the heart. Cardiac surgeons also suggested superimposing the CT information on the smart glasses or conventional loupes, which they referred to as ‘smart loupes’.

Active constraints

Despite orthopaedic surgeons initially stated that they do not need ACs, further discussions with the panel of expert orthopaedic surgeons, who also have some experience with robotics and MIS, at SMARTsurg 1st PC meeting in Milan, concluded that ACs are needed and could be useful to prevent injuries to the rim of the meniscus. Moreover, AC could be used to minimize cutting of the meniscus during surgery. ‘Parrot beak tear’ and ‘Flap tear’ are exceptional cases, where the ACs could help just to remove the flaps. Moreover, in the case of ‘Bucket Handle Tear’, ACs could be helpful to restrict the movement of the instrument in the red zone of meniscus where the success of the repair is very high. So, in this case, ACs could be helpful to prevent injury in the red-white and white zone. It could also be helpful to prevent injury to the peroneal nerve during the cauterization for meniscectomy. For urologist, the AC is useful during the lymphadenectomy step of radical prostatectomy to prevent injuries to arteries, veins and nerves or to prevent injury to accessory vessels coming from the pelvic wall side. It could also be useful for the nerve-sparing in RARP. In RAPN, it could be helpful to prevent injury to vena cava and aorta. However, many urologists believe that ACs should only be implemented for surgical training and for junior surgeons. They also need the overriding functionality as they think that it could be a distraction, confusing and may increase the surgery time. For cardiac surgery use cases, ACs would help avoiding many vital structures, for example, vessels, nerves and so on, involved in the surgery. During CABG, in harvesting the left internal mammary artery (LIMA), surgeons have to be cautious not to get too close to the LIMA while cauterizing, where the ACs would be useful.

Articulated instruments

Due to the small area of the complex knee anatomy, orthopaedic surgeons need articulated instruments for suturing of the meniscus tear as well as for visualizing the damaged structures in 3D. For urological use cases, articulated instruments may be especially helpful for RARP. There are structures in the pelvis, very small and in close proximity, for example, ridges of the pubic bone. In complex cases where there are adhesions in pelvis or abdomen, movement of the instruments in the pelvis is difficult due to the peculiar shape of the pubic bone, and the surgery is performed in the narrow area between the prostate and the rectum. Surgeons need articulated instruments to obviate frequent change of ports and gain better access to anatomic sites. The current cardiovascular surgery instruments do not provide 360° rotational movements. During cardiovascular surgery, it is difficult to access some anatomical structures, for example, the access to the heart is provided from the anterior side, while the MV is on the posterior side. Articulated instruments could be helpful to access the ventricles behind the MV and for cross-clamping of the aorta during retrograde cardioplegia.

Master system: Hand exoskeleton

Concerning the hand exoskeleton, surgeons need them for both hands, and they welcome the possibility of using all fingers instead of the index–thumb grip of the current system. The hand exoskeleton should be lightweight and adjustable for different hand sizes with an accompanying armrest. The design of the exoskeleton needs to take into account the limitations of wrist movements and how this can be transferred to the surgical instrument. Specifically, in cardiac surgery use cases, the ability of the exoskeleton tracking a three-finger pencil grip is highly essential. Surgeons also need haptic feedback on the hand exoskeleton for assessing forces which they apply in various surgical tasks.

Limitations of this study and conclusions

Although the elicited requirements were comprehensive, the limitation of this study is the underlying inequality in the number of surgeons interviewed in different surgical specialties and their gender imbalance. However, during the ‘across-case’ analysis, prioritization and analysis have been made on a common set of requirements between the specialties, so the elicited requirements have wide applicability. Moreover, applicability could come from the already wide application base, for example, seven different use cases from orthopaedics, urology and cardiovascular surgery, thus the elicited requirements could be generalized to a broader range of surgical systems. The proposed framework from eliciting requirements could be an interesting approach to obtain a common set of user requirements in a development of the cross-disciplined robotic systems considering technologies and interfaces. Although the interviews were

conducted within multi-specialty and at different centres, other health-care professionals, for example, nurses or anaesthetists from different countries were not included, which may further elicit new requirements, for example, requirements of new technologies in the preoperative phase.

Author's note

Elena De Momi and Giancarlo Ferrigno are now affiliated with Department of Electronics, Information and Bioengineering (DEIB), Politecnico di Milano, Milan, Italy.


Declaration of conflicting interests

The author(s) declared no potential conflicts of interest with respect to the research, authorship, and/or publication of this article.

Funding

The author(s) disclosed receipt of the following financial support for the research, authorship, and/or publication of this article: This project has received funding from the European Union's Horizon 2020 research and innovation programme under grant agreement No. H2020-ICT-2016-732515.

ORCID iD

Hirenkumar Nakawala  <https://orcid.org/0000-0002-0121-2828>

Supplemental material

Supplemental material for this article is available online.

References


- Hoeckelmann M, Rudas IJ, Fiorini P, et al. Current capabilities and development potential in surgical robotics. *Int J Adv Robot Syst* 2015; 12(5): 61.
- Tan A, Ashrafian H, Scott AJ, et al. Robotic surgery: disruptive innovation or unfulfilled promise? A systemic review and meta-analysis of the first 30 years. *Surg Endosc* 2016; 30(10): 4330–4352.
- BenMessaoud C, Kharrazi H, and MacDorman KF. Facilitators and barriers to adopting robotic-assisted surgery: contextualizing the unified theory of acceptance and use of technology. *PLoS One* 2011; 6(1): e16395.
- Alemzadeh H, Raman J, Leveson N, et al. Vents in robotic surgery: a retrospective study of 14 years of FDA data. *PLoS One* 2016; 11(4): e0151470.
- Tzemanaki A, Walters P, Pipe AG, et al. An anthropomorphic design for a minimally invasive surgical system based on a survey of surgical technologies, techniques and training. *Int J Med Robot* 2014; 10(3): 368–378.
- Smith AB, Raynor M, Amling CL, et al. Multi-institutional analysis of robotic radical cystectomy for bladder cancer: perioperative outcomes and complications in 227 patients. *J Laparoendoscopic Adv Surg Tech* 2012; 22(1): 17–21.
- Herron DM and Marohn M. The SAGES-MIRA robotic surgery consensus group. A consensus document on robotic surgery. *Surg Endosc* 2008; 22(2): 313–325.
- SAFROS Project: patient safety in robotic surgery. Patient safety, World Health Organization (WHO), <http://www.who.int/patientsafety/safesurgery/safros/en/> (2017, accessed 10 December 2017).
- Randell R, Alvarado N, Honey S, et al. Impact of robotic surgery on decision making: perspectives of surgical teams. *AMIA Annu Symp Proc* 2015; 2015: 1057–1066.
- Bonfè M, Boriero F, Dodi R, et al. Towards Automated surgical robotics: a requirements engineering approach. In: *Proceedings of IEEE international conference on biomedical robotics and biomechanics*, Rome, 24–27 June 2012, pp. 56–61.
- Stollnberger G, Moser C, Giuliani M, et al. User requirements for a medical robotic system: enabling doctors to remotely conduct ultrasonography and physical examination. In: *Proceedings of the IEEE international symposium on robot and human interactive communication (RO-MAN)*, New York, 26–31 August 2016, pp. 1156–1161.
- Titan Medical finalizes user requirements for 1st generation robotic surgical system; reiterates 2017 and 2018 development milestones leading to FDA 510(k) submission, Press Releases, Titan Medical (Canada), <https://titanmedicalinc.com/titan-medical-finalizes-user-requirements-for-1st-generation-robotic-surgical-system-reiterates-2017-and-2018-development-milestones-leading-to-fda-510k-submission/> (accessed 13 October 2017).
- Davis B. Robotic surgery – a personal view of the past, present and future. *Int J Adv Robot Sys* 2015; 12(5): 54.
- Aaltonen IE and Wahlström M. Envisioning robotic surgery: surgeons' needs and views on interacting with future technologies and interfaces. *Int J Med Robot* 2018; 14: e1941.
- Zowghi D and Coulin C. Requirements elicitation: a survey of techniques, approaches, and tools. In: A. Aurum and C. Wohlin (eds) *Engineering and managing software requirements*, Springer, Berlin, Heidelberg, 2005, pp. 19–46.
- Tzemanaki A, Burton TM, Gillatt D, et al. μ Angelo: a novel minimally invasive surgical system based on an anthropomorphic design. In: *The 5th IEEE RAS & EMBS international conference on biomedical robotics and biomechanics*, Sao Paulo, Brazil, 12–15 August 2014, pp. 369–374.
- Ayres L, Kavanaugh K, and Knafl KA. Within-case and across-case approaches to qualitative data analysis. *Qual Health Res* 2003; 13(6): 871–883.
- NHS Workforce Statistics, NHS Digital. <https://digital.nhs.uk/data-and-information/find-data-and-publications/supplementary-information/2018-supplementary-information-files/hchs-drs-by-specialty.-grade-and-gender-april-2015-april-2018> (accessed 5 May 2018).
- Nemoto K, Suzuki Y, and Kondo Y. Current work environments: What problems are being faced by Japanese urologists? *Int J Urol* 2018; 25(4): 327–336.
- Sharan M and Elisabeth T. *Qualitative research: A guide to design and implementation*. 4th ed. Hoboken: Jossey-Bass, John Wiley & Sons, 2016.
- Strauss AL and Corbin JM. *Basics of qualitative research: grounding theory procedures and techniques*. 2nd ed. Thousand Oaks: Sage, 1990.

Appendix I


Use cases workflow

This appendix represents the surgical workflow of surgical use cases, mentioned in section ‘Use cases’, in a graphical format. The workflows have been specified using three components of the surgical activity: (1) Phases and their precedence are specified in the top row of the graph with ‘start’ and ‘end’ markings; (2) the middle row shows a sequence of surgical steps for each phase; and (3) the last row specifies the instruments used in the individual phases of each workflow. RaPLM, RaLMR, CABG and MV Surgery are not currently robot-assisted, so the name refers to the goal of the SMARTsurg project demonstrator.


1. Robot-assisted Partial Lateral Meniscectomy (RaPLM)




2. Robot-assisted Repair of Lateral Meniscus Tear (RaLMR)





3. Robot-assisted partial nephrectomy (RAPN)




4. Robot-assisted radical prostatectomy (RARP)










5. Robot-assisted cystectomy and intracorporeal reconstruction with ileal conduit or orthotopic neobladder (RARC)







6. Robot-assisted coronary artery bypass grafting (CABG)





7. Robot-assisted Mitral Valve surgery (MV surgery)





Appendix 2

'Within-case' analysis

Table 2A. Orthopaedics surgery: 'Within-case' analysis.

| No. | Questionnaire |
|-----|--|
| 1 | <p>What are the barriers of current methods that you use (open surgery/manual MIS/RAMIS) in terms of:</p> <p>1. Vision?</p> <p>2. Instruments (slave system: instruments and robotic arms)?</p> <p>3. Interface (master system that the surgeon uses)?</p> <p>Vision Open surgery: There are no barriers. MIS: To see the knee compartments, the camera ports require changing frequently. Assistants also have to change the knee positions repeatedly. With the current camera, it is difficult to see anatomy, for example, to see at the back of meniscus, and in the case of soft tissue obstruction.</p> <p>Instruments RAMIS: – Open surgery: Current technique for meniscus damage, that is, using the probe, is not very efficient. MIS: Instruments' size is big and needs to be miniaturised.</p> <p>Interface RAMIS: – Open surgery: Surgeon's position is not ergonomic. MIS: There is no haptic feedback and surgeon's position is not ergonomic.</p> |
| 2 | <p>What affects your surgical resilience during long procedures?</p> <p>Frequently changing the knee positions, handling instruments in MIS, complexity of surgery and inexperienced assistants affect the surgical resilience during long procedures.</p> |
| 3 | <p>What feature(s) do you not have in manual MIS that you have in open surgery and that you wish you had?</p> <p>Do you find the manipulation of tissues using MIS instruments restrictive as compared to your own hand?</p> <p>MIS lacks the haptic feeling. The manipulation of tissues using MIS instruments is restrictive.</p> |

(continued)

Table 2A. (continued)

| No. | Questionnaire |
|-----|--|
| 4 | What kind of grasps do you use during open/MIS/RAMIS? What different grasping methods/grasping instruments would you welcome? There are three types of grasper, that is, cutter (17%), suture (16%), arthroscopic graspers (67%), reported by the surgeons. The grasping is not sufficient and the grasped tissue is often lost, so more stronger jaw force would be welcomed. |
| 5 | What would you change about current MIS instruments? The surgeons prefer to have small articulated instruments for performing surgery through medial meniscus posterior horn and for stitching the meniscus tear. Surgeons would also like to have the haptic feedback. |
| 6 | Would a third finger be of use? Yes. For example, third finger may be useful to stabilise the meniscus and other fingers could be used to cut it. It could also be useful for meniscus repair and suturing. |
| 7 | Would you want the instruments to have tips that can be swapped over so that the same main instrument can perform as different tools if it has more than one digits? Yes. It could be useful to reduce the infection for instance. |
| 8 | How would you prefer to control the instruments? Using teleoperation? What kind of interface? At the moment, the instruments are controlled manually both in open and laparoscopic surgery. If using the teleoperation, surgeons prefer 'Hand exoskeleton' (43%), followed by 'Omni phantom' (15%), 'da Vinci master console' (14%), 'Cyberglove' (14%), 'Leap motion' (7%) and 'None of these' (7%). |
| 9 | Do you use cameras/endoscopes/laparoscopes? Yes, the surgeons use arthroscopes. |
| 10 | What are your requirements in terms of field of view? The field of view (2D) requirement is less than 1 cm ² to 4 cm ² . 2 cm ² viewing area is sufficient to visualise the whole knee compartment. Larger field of view is helpful to identify the parts of meniscus, to avoid complication, for example, damage to peroneal nerve. |
| 11 | Do you need visual feedback in wider areas, for example, behind obstacles (other organs)? Yes (67%), No (33%). Visual feedback is needed during the meniscus repair to put the suture through the meniscus and to feel correct length of the thread. It is also needed to see the suture and its correct position, for example, start and end position and prevent injury to arteries, for example, popliteal artery. |
| 12 | When operating, do you communicate efficiently with the rest of the surgical team? Yes (75%), No (25%) |
| 13 | In respect to visual feedback, would you welcome such information displayed in your vision during surgery? If yes, what kind of information (e.g. physiological data)? Yes (40%), No (60%). Surgeons would like to see pre-operative images, for example, MRI, blood pressure and other vital signs. Immersive stereo viewer is helpful and smart glass for assistants may be helpful. |
| 14 | Is a teleoperated camera holder required? Yes (80%), No (20%) |
| 15 | How would you prefer the camera was controlled (e.g. voice commands, eye-gaze tracking, head movements, foot pedal, other)? By 'Head movements' (29%), 'Voice control' (28%), 'Eye-gaze tracking' (0%), 'Pedal' (0%), 'None of these' (43%) |
| 16 | Would you wish to move, extend or focus the field of view by moving your head around? No. End-users prefer smart glasses |
| 17 | How could 'active constraints' help you during a surgical operation? Would you like knowing that the instrument would not enter or even touch the boundaries of forbidden regions and/or tissues labelled by you (the surgeon) in a pre-operative and operative stage? Would you like the robot to keep the instrument at a certain angle, for example, normal to the operating path, specified by you to help you guide it? Yes (17%), No (83%). Surgeons found it helpful to avoid damage to the cartilages while doing meniscus repair. However, surgeons wanted an overriding capability to this functionality. Regarding keeping the instrument at certain angle, surgeons commented that there is no enough space to implement the active constraints. |
| 18 | a. How important is haptic feedback during surgery for you? b. What type of haptic feedback would be useful to you (e.g. force feedback of pulling/pushing tissue and surrounding structures or of the thread tension during suturing, force feedback during grasping, texture, temperature?) Yes, it is very important. It is useful for force feedback of pulling/pushing tissue and surrounding structures or of the thread tension during suturing. c. Would it be helpful to 'exaggerate' this feeling, that is, scaled up from the measured exerted force on the tissue? Yes |

(continued)

Table 2A. (continued)

| No. | Questionnaire |
|-----|---|
| | d. Would alternate sensory information be useful as a replacement to haptic feedback or as complimentary to it (e.g. acoustic signals/visual cues/vibration proportional to the exerted force on the tissue or as alarm for over-the-threshold forces)? 'Visual cues' (50%), 'Acoustic signals' (25%), 'Vibration' (25%), 'No alternative sensation' (0%) |
| 19 | a. Do you use pre-operative images? If yes, what type and why? Magnetic Resonance Imaging (MRI) and X-ray are used to assess if there is the meniscus tear. Yes (40%), No (40%), only if 3D (20%) b. When would you need to superimpose such images on the vision of the laparoscope (e.g. to guide/help you identify structures in the abdomen?) Superimposing information is needed, where landmarks, for example, medial condyle of tibia, are useful to register and identify the structures. c. How different is the operating field from the pre-op images (e.g. in terms of tissue deformation?) 'Major difference' (80%), 'No difference' (0%), 'Sometimes different' (20%) |
| 20 | How do you expect a system like SMARTsurg will improve in new surgeon's training? If the system would implement the virtual reality operations, and presents different scenarios then it could be helpful as a simulator, like a video game, for surgical training. However, surgeons prefer same existing training paradigm and found some components would be limited, for example, active constraints. |
| 21 | Any other concerns about the technology? Surgeons were concerned about increased cost of the procedure. The surgeons were also concerned with the manipulation of tissues with teleoperation while using the robots especially with the knee joint because the space is very confined. |

Table 2B. Urological surgery: 'Within-case' analysis.

| No. | Questionnaire |
|-------------|--|
| I | What are the barriers of current methods that you use (open surgery/manual MIS/RAMIS) in terms of: 1. Vision? 2. Instruments (slave system: instruments and robotic arms)? 3. Interface (master system that the surgeon uses)? |
| Vision | Open surgery: There are problems visualising small anatomical structures with the loupes in the pelvis, for example, anterior part of the prostate – the apex, urethra, venous plexus and the cleavage between the prostate and rectum. MIS: There are problems with the field of view, for example, in coordination with the vision, since the vision is unidirectional. There is poor visibility when there are adhesions in the pelvis. RAMIS: Camera needs frequent cleaning and requires keeping it close to the structures. There are problems accessing some anatomy, for example, ridges of public bone. |
| Instruments | Open surgery: The instruments are not flexible. It is not easy to reach the areas in pelvis, especially in radical prostatectomy. MIS: The coordination of action and vision is difficult. Moreover, the instruments are not flexible. RAMIS: The instruments are not flexible and force feedback is missing. Retraction of tissues by assistants are missing. The current instruments of tissue retraction are smaller. There are possibility to collateral damage due to repeated change of instruments. Some procedures, for example, bowel anastomosis and cold cutting, are difficult with da Vinci instruments. |
| Interface | Open surgery: – MIS: – RAMIS: Surgeon's position is not ergonomic. There is no back rest. The master controller is not very efficient since it requires frequent clutching and the arms collide with each other many times. There is no tactile feedback. |
| 2 | What affects your surgical resilience during long procedures? Surgeon's current sitting position with dVSS, instrument manipulation in confined spaces in pelvis, 3D vision, which was causing tiring, pain, and redness to some surgeons, and cognitive load with respect to complexity of surgery were affecting surgical resilience. |
| 3/4/5 | What feature(s) do you not have in RAMIS that you have in open surgery and that you wish you had? If you are a da Vinci user, is there anything specific that you cannot do using the da Vinci surgical system? Please think of examples. What would enable you to tackle this challenge? What feature(s) do you not have in manual MIS that you have in open surgery and that you wish you had? - With RAMIS, there is no haptic feedback, image resolution is poor with zoom. There is also unavailability of bigger forceps and tissue retraction instruments, which are available in open surgery. - The surgeons using dVSS cannot be able to manipulate tissues when there are tissue adhesions and with the close structures inside the pelvis. There is also limitation with the field of view. Superimposing pre-operative information, that is, Computed Tomography (CT) or Magnetic Resonance Imaging (MRI), would helpful to tackle this challenge. - With MIS, there is no haptic feedback. |

(continued)

Table 2B. (continued)

| No. | Questionnaire |
|-----|---|
| 6/7 | Do you find the manipulation of tissues using MIS instruments restrictive as compared to your own hand? In the case of manipulation of tissues using RAMIS instruments restrictive as compared to your own hand? Yes, the surgeons find manipulation of tissues using MIS instruments restrictive because there is no haptic feeling. There are less variety of instruments available, for example, instruments for tissue retraction, which restrict access to inaccessible anatomy otherwise accessible by hands. |
| 8 | What kind of grasps do you use during open/MIS/RAMIS? What different grasping methods/grasping instruments would you welcome? Urologists use different graspers, that is, ProGrasp forceps (25%), Bipolar forceps (21%), Maryland forceps (21%), PK dissecting forceps (9%), Cadiere forceps (8%), Electro-cautery grasper (4%), Bowel grasper (4%), Retractor (4%), Kelly forceps (4%). Surgeons prefers the graspers which could provide stronger grip than the current graspers and require less replacement. |
| 9 | What would you change about current MIS/RAMIS instruments? The current instruments should be articulated and disposable. Grasping mechanism should allow more force while grasping. |
| 10 | Would a third finger be of use? Yes (33%), No (67%). A third finger would be of great use during the mono-port surgeries since it provides more articulation, and it could help during dissection, stabilising tissue, suturing and grasping, for example, to Gerota's fascia. However, surgeons were concerned that it may conflict with the instrument arms. |
| 11 | Would you want the instruments to have tips that can be swapped over so that the same main instrument can perform as different tools if it has more than one digits? Yes (67%), No (33%). For example, during the RAPN, change of monopolar curved scissors to robotic large needle driver. It could also save time and may be helpful during stitching and bleeding. It also allows to work easily with inefficient assistants. |
| 12 | How would you prefer to control the instruments? Using teleoperation? What kind of interface? At the moment, the instruments are controlled manually both in open and laparoscopic surgery. With da Vinci, surgeons prefer using teleoperation. Other accepted interfaces are 'Hand exoskeleton' (40%), 'Cyberglove' (28%), 'da Vinci master console' (20%), 'Leap Motion' (8%), 'Omni phantom' (4%), 'None of these' (0%). |
| 13 | Do you use cameras/endoscopes/laparoscopes? Urologists use laparoscopes. |
| 14 | Are they 2D/3D? They are 2D and 3D. |
| 15 | What are the barriers in the laparoscope of the da Vinci/laparoscope and how do you think they could be overcome? Image resolution is poor. Camera is small and gets often dirty. The camera is also inflexible that limits its accessibility to unreachable regions, that is, in prostatectomy. da Vinci's camera length is around 30 cm which often clashes with assistant instruments. |
| 16 | What are your requirements in terms of field of view? The requirements in terms of field of view is 5 cm ² to 25 cm ² . |
| 17 | Do you need visual feedback in wider areas, for example, behind obstacles (other organs)? Yes (80%), No (20%). Visual feedback in the wider areas are helpful in certain conditions, for example, to locate bowel or long structures, to see big vessels, renal vein and arteries behind fat, and tumour nodes, and to look behind the obstacles, for example, during anastomosis in radical prostatectomy. |
| 18 | When operating, do you communicate efficiently with the rest of the surgical team? Yes (80%), No (20%) |
| 19 | If you are a da Vinci user, do you feel immersed in the da Vinci console? If yes, do you welcome this or would you prefer to also have greater awareness of your surrounding environment? While most of the urologists think, the da Vinci is immersive, they think immersive stereo viewer could be an alternative in the new system, while the smart glasses could be used by the assistant surgeons. |
| 20 | In respect to visual feedback, would you welcome such information displayed in your vision during surgery? If yes, what kind of information (e.g. physiological data)? Yes (18%), No (82%). Urologists would like to see intra-abdominal pressure and information on the blood loss. |
| 21 | If you are a da Vinci user, how would you rate the da Vinci's system in terms of efficiency and ergonomics? End-users think that da Vinci's efficiency and ergonomics are very good. |
| 22 | Is a teleoperated camera holder required? Yes |
| 23 | How would you prefer the camera was controlled (e.g. voice commands, eye-gaze tracking, head movements, foot pedal, other)? 'Head movements' (33%), 'Pedal' (34%), 'Eye-gaze tracking' (22%), 'Voice control' (0%), 'None of these' (11%) |

(continued)

Table 2B. (continued)

| No. | Questionnaire |
|-----|--|
| 24 | Would you wish to move, extend or focus the field of view by moving your head around? Yes (with the immersive stereo viewer) |
| 25 | How could 'active constraints' help you during a surgical operation? Would you like knowing that the instrument would not enter or even touch the boundaries of forbidden regions and/or tissues labelled by you (the surgeon) in a pre-operative and operative stage? Would you like the robot to keep the instrument at a certain angle, for example, normal to the operating path, specified by you to help you guide it? Yes (45%), No (33%), only for training (e.g. on simulator and for young surgeons) (22%). Urologists think that the AC may be useful for preventing injuries to big vessels, for example, vena cava, renal arteries during the kidney surgery. During prostatectomy, AC could be helpful preventing injuries to small vessels, for example, accessory vessels and small nerves during lymphadenectomy step. Vena cava, aorta, rectum could be labelled as forbidden regions/'no-go' zones. While there was mix opinion about overriding facility, most of surgeons disagreed with the robot adjusting instrument angles where the free movement of instruments was preferred. |
| 26 | a. How important is haptic feedback during surgery for you? b. What type of haptic feedback would be useful to you (e.g. force feedback of pulling/pushing tissue and surrounding structures or of the thread tension during suturing, force feedback during grasping, texture, temperature?) Yes (69%), No (8%), only for training (23%). All type of feedback is welcomed. It could be helpful with the training, for example, during the learning curve to identify the pubic bone. Haptic feedback could be useful to suturing the parenchyma in kidney surgery, dissecting organs or to identify the remaining tumour. c. Would it be helpful to 'exaggerate' this feeling, that is, scaled up from the measured exerted force on the tissue? Yes (27%), No (73%) d. Would alternate sensory information be useful as a replacement to haptic feedback or as complimentary to it (e.g. acoustic signals/visual cues/vibration proportional to the exerted force on the tissue or as alarm for over-the-threshold forces)? 'Visual cues' (35%), 'Vibration' (18%), 'Acoustic signals' (12%), 'Combined (visual cues and vibration)' (12%), 'No alternative feedback' (23%) |
| 27 | a. Do you use pre-operative images? If yes, what type and why? Yes. Mostly CT, MRI, Ultrasonography (USG). For the kidney, CT scan is used. MRI scan, especially multi-barometric MRI, and histopathology are used for prostate and bladder respectively. b. When would you need to superimpose such images on the vision of the laparoscope (e.g. to guide/help you identify structures in the abdomen?) Superimposing images are needed to identify structures, know relative position of structures. There are enough landmarks available, for example, apex of the prostate. c. How different is the operating field from the pre-op images (e.g. in terms of tissue deformation?) It is always different except for the parenchymal organs like kidney |
| 28 | How do you expect a system like SMARTsurg will improve in new surgeon's training? Yes (90%), No (10%). It could be helpful for young surgeons with the simulator or dry lab. The assistants should have 3D glasses for the training and surgeons use the da Vinci. Surgeons also think that it could help with the learning curve. |
| 29 | Any other concerns about the technology? The new technologies should be user friendly and allow easy surgical planning. The size of robot and cost should be reduced. There should be synchronised movements of surgical table and the slave systems, new mechanisms for clutching the master controller and the instruments. |

Table 2C. Cardiovascular surgeries: 'Within-case' analysis.

| No. | Questionnaire |
|--------|--|
| I | What are the barriers of current methods that you use (open surgery/manual MIS/RAMIS) in terms of: 1. Vision? 2. Instruments (slave system: instruments and robotic arms)? 3. Interface (master system that the surgeon uses)? |
| Vision | Open surgery: There are limitations with anatomical structures, for example, in mitral valve surgery, the access is anterior, while the valve is on the posterior side. Surgeons need to see inside the ventricles behind the mitral valve, for example, to replace the chordae. These anatomical structures, along with papillary muscles, thorax and ribs are difficult to visualise. MIS: Same as 'Open surgery'. RAMIS: It is difficult to perceive the depth. The current size of the da Vinci's camera port is larger, for example, 8 mm. |

(continued)

Table 2C. (continued)

| No. | Questionnaire |
|-------------|--|
| Instruments | <p>Open surgery: –</p> <p>MIS: Fulcrum effect limits the movements. Physical access is limited and hand–eye coordination is not optimal. There is no haptic feeling. Instruments do not provide 360° rotational movements. It is difficult to move the full arm.</p> <p>RAMIS: –</p> |
| Interface | <p>Open surgery: There are limitations of wrist movements even in open surgery. The anatomical structures are in the awkward position and sometimes it is needed to move the patient body to adjust the angles.</p> <p>MIS: –</p> <p>RAMIS: –</p> |
| 2 | <p>What affects your surgical resilience during long procedures?</p> <ul style="list-style-type: none"> - Long time in standing position, for example, surgeons generally stand in the area of 40 cm² for minimum of 2 h and maximum 5 h with wearing all the things constantly, for example, conventional loupes. - No arm rests. - In MIS, vision is adjusted by the assistant and the arms needs to be adjusted by the surgeons, and this causes the tiredness in long procedures. - In less invasive surgery, due to keyhole surgery, limited instruments movements, repeated actions, limited vision and haptic feeling reduces the concentration and increases the learning curve. With the open surgery, if there is a complication, it is tiring because it increases the surgery time. |
| 3 | <p>What feature(s) do you not have in manual MIS that you have in open surgery and that you wish you had?</p> <p>Do you find the manipulation of tissues using MIS instruments restrictive as compared to your own hand?</p> <ul style="list-style-type: none"> - The surgeons would like to adjust the camera by themselves. There are also less angles available than the open surgery. It is also difficult to deliver retrograde cardioplegia in MIS because it is hard to cross clamp the aorta in MIS. The camera also gets dirty often in MIS. The vision is good with open surgery and would like to have in MIS. - Manipulation of tissues using MIS instruments is less informative because the instruments are longer and manipulation is indirect. It is easier to manipulate in open surgery due to haptic feeling. RAMIS instruments are as good as hands but instruments need the concept of the pencil grip mechanism. |
| 4 | <p>What kind of grasps do you use during open/MIS/RAMIS? What different grasping methods/grasping instruments would you welcome?</p> <p>Surgeons use different type of graspers, that is, traditional graspers (33%), Resano forceps (17%), Debakey forceps (17%), Coronary forceps (17%), Pencil grip instruments for forceps and needle holders (16%). It would be great to have a pencil grip like grasping mechanism in MIS/RAMIS.</p> |
| 5 | <p>What would you change about current MIS instruments?</p> <p>Surgeons would like to have improved tactile feedback and the range of movements of instruments. Moreover, the concept of pencil grip for instrument jaw grip should be developed. Mechanical strength of the instruments should be improved during suturing.</p> |
| 6 | <p>Would a third finger be of use?</p> <p>Yes (80%), No (20%). It could be useful for the rotation movements obtained using the pencil grip, for example, Castroviejo-type instrument and to cut the sutures, for example, during the anastomosis.</p> |
| 7 | <p>Would you want the instruments to have tips that can be swapped over so that the same main instrument can perform as different tools if it has more than one digits?</p> <p>Yes. Especially, it could be useful for cutting the sutures that are required to cut by assistants.</p> |
| 8 | <p>How would you prefer to control the instruments? Using teleoperation? What kind of interface?</p> <p>At the moment, the instruments are controlled manually both in open and laparoscopic surgery.</p> <p>‘Hand exoskeleton’ (46%), ‘Leap Motion’ (15%), ‘Cyberglove’ (15%), ‘da Vinci master console’ (8%), ‘Nintendo wii’ (8%), ‘Omni phantom’ (8%)</p> |
| 9 | <p>Do you use cameras/endoscopes/laparoscopes?</p> <p>Cardiac surgeons use loupes with magnification lenses and endoscopes.</p> |
| 10 | <p>What are your requirements in terms of field of view?</p> <p>The requirements in terms of (2D) field of view is less than 1.5 mm² to 7 cm².</p> |
| 11 | <p>Do you need visual feedback in wider areas, for example, behind obstacles (other organs)?</p> <p>Yes (60%), No (40%). Visual feedback would remove the need of an assistant to control your vision.</p> |
| 12 | <p>When operating, do you communicate efficiently with the rest of the surgical team? Would you prefer to also have greater awareness of your surrounding environment?</p> <p>Yes. The surgeons prefer immersive stereo viewer. There are specific protocols for communications in OR and it is very important.</p> |
| 13 | <p>In respect to visual feedback, would you welcome such information displayed in your vision during surgery? If yes, what kind of information (e.g. physiological data)?</p> |

(continued)

Table 2C. (continued)

| No. | Questionnaire |
|-----|--|
| | Yes, the cardiac surgeons would like to see physiological data, for example, heart rate, blood pressure, oxygen saturation, central venous pressure, electrocardiogram, body temperature. |
| 14 | Is a teleoperated camera holder required? |
| | Yes |
| 15 | How would you prefer the camera was controlled (e.g. voice commands, eye-gaze tracking, head movements, foot pedal, other)? |
| | 'Voice control' (28%), 'Eye-gaze tracking' (27%), 'Pedal' (18%), 'Hand movements' (18%), 'Something else' (9%) |
| 16 | Would you wish to move, extend or focus the field of view by moving your head around? |
| | Yes |
| 17 | How could 'active constraints' help you during a surgical operation? Would you like knowing that the instrument would not enter or even touch the boundaries of forbidden regions and/or tissues labelled by you (the surgeon) in a pre-operative and operative stage? Would you like the robot to keep the instrument at a certain angle, for example, normal to the operating path, specified by you to help you guide it? |
| | Yes (83%), No (17%). AC could be useful to prevent injuries to critical structures, for example, vessels such as LIMA (Left Internal Mammary Artery) during CABG and circumflex branch of left coronary artery during MV surgery. There are no labelling regions for both the use cases, however SA (sinoatrial) node and coronary artery could be used. Cardiac surgeons suggested that AC should be used as alerts. It should not be guiding or adjusting angles of instruments. |
| 18 | a. How important is haptic feedback during surgery for you? |
| | b. What type of haptic feedback would be useful to you (e.g. force feedback of pulling/pushing tissue and surrounding structures or of the thread tension during suturing, force feedback during grasping, texture, temperature?) |
| | Yes, because it is a disadvantage of MIS. For tissue repair and to assess calcium deposits in artery, surgeons need to feel the tissue. Force feedback of pulling/pushing tissue and surrounding structures or of the thread tension would be useful. |
| | c. Would it be helpful to 'exaggerate' this feeling, that is, scaled up from the measured exerted force on the tissue? |
| | Yes (50%), No (33%), only after training (17%) |
| | d. Would alternate sensory information be useful as a replacement to haptic feedback or as complimentary to it (e.g. acoustic signals/visual cues/vibration proportional to the exerted force on the tissue or as alarm for over-the-threshold forces)? |
| | No, it would not be helpful, that is, 'Acoustic signals' (17%), 'vibration' (17%), 'visual cues' (16%), 'No alternative sensation' (50%) |
| 19 | a. Do you use pre-operative images? If yes, what type and why? |
| | Combination of echocardiography, coronary angiography, CT, MRI |
| | b. When would you need to superimpose such images on the vision of the laparoscope (e.g. to guide/help you identify structures in the abdomen?) |
| | Yes (67%), No (16%), only if in trouble (17%) |
| | Superimposing information is needed. There are many landmarks available, for example, appendages, great vessels, the apex which would be helpful to identify the structures. |
| | c. How different is the operating field from the pre-op images (e.g. in terms of tissue deformation?) |
| | 'Major difference' (17%), 'Less difference' (33%), 'No difference' (50%). For example, images are different in the case of degenerative mitral valve and endocarditis. |
| 20 | How do you expect a system like SMARTsurg will improve in new surgeon's training? |
| | Yes (75%), No (25%). It could be an ideal way to use virtual reality, robotic endoscopes and the animals to use for training. Bed-side surgeons with smart glasses could be used for the training. |
| 21 | Any other concerns about the technology? |
| | The surgeons were concerned about patient safety, due to involvement of critical structures, while using the robots for cardiovascular surgery. |

Appendix 3

Table 3A. ‘Across-case’ analysis: Non-mandatory requirements.

| No. | Elicited requirement | Orthopaedics | Urology | Cardiovascular surgery | Total score |
|-----|---|---|--|--|-------------|
| 5 | Image resolution Details | Better image quality is helpful. | Better image resolution, for example, 4K (3840 × 2160 px) – Ultra HD image quality | Good image resolution, at least at the level of conventional loupes, is needed. Surgeons also require magnification (2.5× to 3.5×) and a larger field of view (from 1.5 mm ² to 5 cm ²) | |
| | Scores | 5 | 3 | 5 | 13 |
| 6 | Three-fingered instrument Details | The three-fingered instrument could be helpful to stabilise the meniscus in meniscus repair. It could also be useful to view knee compartments, to cut free cartilage pieces, and to repair tendons and nerves. | Urologists do not need the three-fingered instrument. However, they would like to try first if implemented. The wrist articulation is missing in three-fingered instrument and it may not provide same articulation as the da Vinci single port. | The three-fingered instrument is helpful for cutting the sutures. The cardiac surgeons also showed a willingness to try on a prototype. | |
| | Scores | 4 | 4 | 5 | 13 |
| 7 | Smart glasses Details | Smart glasses are required for assistants and for surgical training. | Smart glasses are required for assistants and for surgical training. | Smart glasses are required for assistants and surgical training. | |
| | Scores | 5 | 3 | 4 | 12 |
| 8 | Haptics Details | Haptics could be required during the tissue manipulation. | Haptics could be required to feel the planes between prostate and rectum, to feel pushing and pulling tissues, thread tension during suturing. It is also helpful for surgical training. | Haptics could be useful to feel the calcium deposit in the coronary artery. | |
| | Scores | 3 | 3 | 4 | 10 |
| 9 | 3D images Details | | 3D images are helpful for better visualisation. | 3D images are helpful for better visualisation. | |
| | Scores | – | 5 | 5 | 10 |
| 10 | Alternative haptic sensation Details | As an alternative haptic feedback, visual cues could be helpful. | As an alternative haptic feedback, visual cues could be helpful. | Cardiac surgeons do not prefer to have an alternative haptic sensation, but they only prefer the naturalistic feedback. | |
| | Scores | 3 | 5 | 1 | 9 |
| 11 | Extended visual feedback Details | Extended visual feedback is helpful to put the suture through the meniscus and to feel the correct length; to see popliteal artery, where more narrow or flexible camera is also needed. | Extended visual feedback is helpful in radical prostatectomy or trans-corporeal reconstruction to see big vessels, renal arteries behind fat and lymph nodes near vena cava or aorta. | Extended visual feedback is helpful to see the critical structures in heart, for example, aorta. | |
| | Scores | 1 | 5 | 3 | 9 |

(continued)

Table 3A. (continued)

| No. | Elicited requirement | Orthopaedics | Urology | Cardiovascular surgery | Total score |
|-----|--|--|---|---|-------------|
| 12 | Flexible camera Details | | Flexible camera is required to look around corners. For example, it could be used similar way as automatic flexible cystectomy, flexible sigmoidoscopy or colonoscopy. | Flexible camera is required to reach or visualise some anatomical structures, for example, mitral valve, ventricles behind the mitral valve, and during the cross-clamping of aorta. It should be designed like a bronchoscope. | |
| | Scores | – | 5 | 3 | 8 |
| 13 | Needle holder (slave side) Details | A new needle holder for suturing is required in meniscus repair. | New instruments or modification to existing instruments is needed, for example, small needle drivers | | |
| | Scores | 5 | 3 | – | 8 |
| 14 | Instrumentation Details | A new needle holder for suturing is required in meniscus repair. | New instruments or modification to existing instruments is needed. For example, thin instruments and small needle drivers are useful aiding minimally invasiveness. Bigger instruments to handle bowel (larger jaws and more force on tissue), to take specimen out and retractors for tissue retraction would be helpful. An easier system to put clips, for example, Hem-o-lok clips and disposable instruments are required. | | |
| | Scores | 5 | 3 | – | 8 |
| 15 | Immersive stereo viewer Details | | An immersion stereo viewer is required for better immersive experience. It should be used by surgeons. | An immersion stereo viewer is required for better immersive experience. It should be used by surgeons. | |
| | Scores | – | 3 | 4 | 7 |
| 16 | Camera control Details | Orthopaedic surgeons prefer using joystick or exoskeleton or hand control. | Head movements are required to control the camera. | Cardiovascular surgeons require voice control, specifically in the big field, and another finer control in the focused field. | |
| | Scores | 3 | 1 | 3 | 7 |
| 17 | Physiological data Details | Orthopaedic surgeons do not need to see the physiological data. | Urologists do not need to see the physiological data. | The physiological data, for example, vital signs which needs to be verify repeatedly, is essential in cardiac surgery. | |
| | Scores | 1 | 1 | 4 | 6 |

(continued)

Table 3A. (continued)

| No. | Elicited requirement | Orthopaedics | Urology | Cardiovascular surgery | Total score |
|-----|---|---|---|--|-------------|
| 18 | Small instruments Details | During RaPLM and RaLMR, knee positions and camera ports are generally required to change repeatedly for better visualisation and manipulation. There are also tissue problems, for example, thin meniscus, where the smaller instruments, at least size of current instrument, for example, 4 mm, are would help for doing surgery through medial meniscus posterior horn for stitching of meniscus tear. | | | |
| | Scores | 5 | — | — | 5 |
| 19 | Teleoperation Details | Teleoperation is required for minimal meniscus resection because surgeons' posture is not good during RaPLM and RaLMR. | | | |
| | Scores | 5 | — | — | 5 |
| 20 | Instrument jaw grip Details | | | Instruments which could provide pencil grip-like mechanism are required. | |
| | Scores | — | — | 5 | 5 |
| 21 | Magnified vision Details | | Magnified vision is helpful for better visualisation | | |
| | Scores | — | 5 | — | 5 |
| 22 | Magnified haptic/ force feeling Details | The exaggerated haptic feeling is required to reduce iatrogenic complications. | Realistic haptic feedback is required. | For clinical purposes, it should not be magnified, but kept within the physiological ranges. | |
| | Scores | 1 | 3 | 1 | 5 |
| 23 | Surgeon's position Details | Ergonomic surgeon's position is required. | Ergonomics and surgeon's position should be considered when constructing the actual master interface. | For cardiovascular applications, resting the forearms while operating with fingertips is important to minimise tremor. | |
| | Scores | 1 | 1 | 3 | 5 |
| 24 | Camera size Details | | | Due to smaller access to the operating area, the size of camera should be less. | |
| | Scores | — | — | 4 | 4 |
| 25 | Teleoperated vision system Details | | | Teleoperated vision system is helpful to remove camera handling by assistants. | |
| | Scores | — | — | 3 | 3 |

(continued)

Table 3A. (continued)

| No. | Elicited requirement | Orthopaedics | Urology | Cardiovascular surgery | Total score |
|-----|--|---|--|--|-------------|
| 26 | Master interface size Details | | | The small interface is desired (35–40 cm ²). | |
| | Scores | – | – | 3 | 3 |
| 27 | Instrument tip swapping Details | Instrument tip swapping is helpful. | Instrument tip swapping is helpful because it remove the need of changing the instruments. | Instrument tip swapping is helpful. | |
| | Scores | 1 | 1 | 1 | 3 |
| 28 | Manipulation with left-handed surgeon Details | Modification to current instruments is needed for left-handed surgeons. | | | |
| | Scores | 1 | – | – | 1 |
| 29 | Field of view Details | | Wider field of view may be helpful to see the assistants' instruments and remove the need of changing the ports. Field of vision should be as larger as possible or needs to be able to 'look around corners'. | | |
| | Scores | – | 1 | – | 1 |
| 30 | Clutching mechanism Details | | New clutching mechanism is needed because frequent clutching is required to handle the workspace limitation. | | |
| | Scores | – | 1 | – | 1 |
| 31 | Easier understanding of surgical workflow steps Details | | It is difficult to handle complex surgical cases and follow the open surgery approaches by the junior surgeons, where explicit procedural guidance may be helpful. | | |
| | Scores | – | 1 | – | 1 |
| 32 | Tissue grasping mechanism Details | | Urologists require newer grasper which could provide more force to grasp tissues. | | |
| | Scores | – | 1 | – | 1 |
| 33 | Length of camera shaft Details | | Short camera length may be helpful to stop clashing of instruments with assistants. | | |
| | Scores | – | 1 | – | 1 |